

CERTIFICATE OF ANALYSIS

AIRGAS WEST
ALAT-AIRGAS WEST
3737 WORSHAM AVE #800
Long Beach, CA 90808
US

Invoice # : 9495818
PO # : 5800084665
Cust. Item # : X05NI79CA3459B2
Certification Date: 30 Mar 2018

Lot Number : 928890
Product : 25PPM H2S/100PPM CO/2.5%CH4/18%O2/N2

Quantity : 3
Product Expiration : Mar 29, 2020
Airgas Cylinder Size: 34DAL

AVERAGE ANALYTICAL CONCENTRATIONS

<u>Component</u>	<u>Concentration</u>	<u>Analytical Uncertainty</u>
CARBON MONOXIDE	100.00 ppm	+/- 2 %
HYDROGEN SULFIDE	25.62 ppm	+/- 5 %
METHANE	2.51 %	+/- 2 %
NITROGEN		
OXYGEN	17.92 %	+/- 2 %

TRACEABILITY

Preparation:


Calibration test 72053, 72057, or 72062 dated, 3rd January 2018 applies.

Weight Sets Used: Kit# 92231, Test# 2565058, Kit#03610, Test#VA-18-10769C, T3 Test# VA-17-10760B

(Gas mixtures manufactured with balance calibrated using N.I.S.T. traceable weights.)

No affecting environmental conditions during analysis.

APPROVED BY:



ALYN SCRITCHFIELD